

Our Ref: [REDACTED] OK

[REDACTED]

Pikua Rense Pty Ltd

PO Box 453  
New Farm 4005

Qld

ATTENTION: C.Tidswell, CEO Pikua Rense Pty Ltd

Received Date	[REDACTED]
Temp Rec'd (°C)	20
Date Started	[REDACTED]
Result Date	[REDACTED]

RE: EVALUATION [REDACTED] HOSPITAL GRADE DISINFECTANT'

Dear [REDACTED]

A sample marked ' [REDACTED] Option B' was evaluated in accordance with the TGA Test for Hospital grade Disinfectants under 'dirty' conditions (Option B). The product was evaluated against the organisms *Staphylococcus aureus* NCTC 4163, *Escherichia coli* NCTC 8196, *Pseudomonas aeruginosa* NCTC 6749, and *Proteus vulgaris* NCTC 4635, in triplicate, utilizing fresh cultures and solutions on each occasion.

#### Test Conditions

<b>Concentration</b>	6g/L Dilution with Hard Water
<b>Contact Time</b>	5 Minutes
<b>Test Temperature</b>	Room Temperature
<b>Validated Neutraliser</b>	T5 10mL

T5 = TSB with 5% Tween, 0.5% Lecithin, and 1% Nutrient Liver Digest

Our Ref: XXXXXXXXXX OK **Results**

Test Number	Organisms	Subculture Number	Inoculum Count CFU/ mL	Ch in Recovery Broths		Pass/ Fail
				1 <sup>st</sup> Challenge	2 <sup>nd</sup> Challenge	
1	<i>S. aureus</i> NCTC 4163	8A	2.85 x 10 <sup>8</sup>	-----	-----	Pass
	<i>E. coli</i> NCTC 8196	8A	5.25 x 10 <sup>8</sup>	-----	-----	Pass
	<i>P. aeruginosa</i> NCTC 6749	8A	5.45 x 10 <sup>8</sup>	-----	-----	Pass
	<i>P. vulgaris</i> NCTC 4635	8A	6.10 x 10 <sup>8</sup>	-----	-----	Pass
2	<i>S. aureus</i> NCTC 4163	8B	2.75 x 10 <sup>8</sup>	-----	-----	Pass
	<i>E. coli</i> NCTC 8196	8B	5.60 x 10 <sup>8</sup>	-----	-----	Pass
	<i>P. aeruginosa</i> NCTC 6749	8B	5.10 x 10 <sup>8</sup>	-----	-----	Pass
	<i>P. vulgaris</i> NCTC 4635	8B	5.70 x 10 <sup>8</sup>	-----	-----	Pass
3	<i>S. aureus</i> NCTC 4163	8C	2.75 x 10 <sup>8</sup>	-----	-----	Pass
	<i>E. coli</i> NCTC 8196	8C	5.55 x 10 <sup>8</sup>	-----	-----	Pass
	<i>P. aeruginosa</i> NCTC 6749	8C	5.00 x 10 <sup>8</sup>	-----	-----	Pass
	<i>P. vulgaris</i> NCTC 4635	8C	6.00 x 10 <sup>8</sup>	-----	-----	Pass

- Notes: 1) CFU = Colony Forming Units  
2) '-' indicates no growth in recovery broths  
3) '+' indicates growth in recovery broths

## Conclusions

The product has **passed** the TGA Option B test under the above test conditions. All controls conformed to the requirements of the test procedure. *Page 3 of 3*

The data pertains solely to the analytical and sampling procedure(s) used and the condition and homogeneity of the sample(s) as received. The data therefore may not **Yours faithfully,**

be representative of the lot or batch or other samples. Consequently the data may not necessarily justify the acceptance or rejection of a lot or batch, a product recall or



support legal proceedings. It is the responsibility of the client to provide all information relevant to the analysis requested. This report does not imply that

similar analysis and for any action that should be taken as a result of the analysis. Australia Pty Ltd has been engaged to consult upon the consequences of the

Imtiaz Ahmed

SCIENTIFIC SERVICES MANAGER

TGA Licence No. MI-04072005-LI-000664-2

